

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091523\
 Data File : VX037759.D
 Acq On : 15 Sep 2023 16:53
 Operator : JC/MD
 Sample : 04390-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10A-GW-150-152

Quant Time: Sep 16 05:46:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	218950	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	362286	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	330839	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	143002	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	131827	45.826	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.660%
35) Dibromofluoromethane	5.385	113	110283	45.495	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.980%
50) Toluene-d8	8.647	98	435948	48.401	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.800%
62) 4-Bromofluorobenzene	11.079	95	149072	45.547	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.100%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	5466	3.950	ug/l #	86
19) Methyl tert-butyl Ether	3.117	73	17415	2.216	ug/l	92
22) Diisopropyl ether	3.764	45	6623	0.894	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091523\
Data File : VX037759.D
Acq On : 15 Sep 2023 16:53
Operator : JC/MD
Sample : 04390-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10A-GW-150-152

Quant Time: Sep 16 05:46:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

