

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091523\
 Data File : VX037762.D
 Acq On : 15 Sep 2023 18:05
 Operator : JC/MD
 Sample : 04459-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE137BASIN-B-20230914

Quant Time: Sep 16 05:47:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	219420	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	364697	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	333699	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	143789	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	133800	46.413	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.820%
35) Dibromofluoromethane	5.391	113	112372	46.050	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.100%
50) Toluene-d8	8.647	98	437702	48.275	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.540%
62) 4-Bromofluorobenzene	11.079	95	148945	45.207	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.420%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	2682	1.211	ug/l	Qvalue 95
16) Acetone	2.380	43	3505	2.527	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091523\
Data File : VX037762.D
Acq On : 15 Sep 2023 18:05
Operator : JC/MD
Sample : 04459-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE137BASIN-B-20230914

Quant Time: Sep 16 05:47:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

