

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091523\
 Data File : VX037763.D
 Acq On : 15 Sep 2023 18:30
 Operator : JC/MD
 Sample : 04459-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE137BASIN-A-20230914

Quant Time: Sep 16 05:48:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	202644	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	334491	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	295880	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117887	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	123495	46.384	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.760%
35) Dibromofluoromethane	5.385	113	103123	46.076	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.160%
50) Toluene-d8	8.647	98	398207	47.885	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.760%
62) 4-Bromofluorobenzene	11.079	95	126696	41.927	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.860%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	2927	1.431	ug/l	Qvalue 93
16) Acetone	2.380	43	4168	3.254	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091523\
Data File : VX037763.D
Acq On : 15 Sep 2023 18:30
Operator : JC/MD
Sample : 04459-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE137BASIN-A-20230914

Quant Time: Sep 16 05:48:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

