

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091619\
 Data File : VX012412.D
 Acq On : 16 Sep 2019 17:55
 Operator : JC/SP
 Sample : K4830-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S13-TB-2019-2

Quant Time: Sep 17 05:14:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 04:36:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	116233	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	190739	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	176951	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	98732	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	125974	52.89	ug/l	0.00
Spiked Amount			Recovery	=	105.78%	
35) Dibromofluoromethane	5.49	113	87772	47.50	ug/l	0.00
Spiked Amount			Recovery	=	95.00%	
50) Toluene-d8	8.71	98	323179	49.29	ug/l	0.00
Spiked Amount			Recovery	=	98.58%	
62) 4-Bromofluorobenzene	11.14	95	127405	46.19	ug/l	0.00
Spiked Amount			Recovery	=	92.38%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091619\
Data File : VX012412.D
Acq On : 16 Sep 2019 17:55
Operator : JC/SP
Sample : K4830-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S13-TB-2019-2

Quant Time: Sep 17 05:14:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091619W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 04:36:38 2019
Response via : Initial Calibration

