


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX091619\
 Data File : VX012401.D
 Acq On : 16 Sep 2019 11:56
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091619W.M
 Title : SW846 8260
 Last Update : Tue Sep 17 04:36:38 2019



AutoFind: Scans 1680, 1681, 1682; Background Corrected with Scan 1673

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	6212	PASS
75	95	30	60	51.2	17006	PASS
95	95	100	100	100.0	33221	PASS
96	95	5	9	6.3	2101	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	71.5	23762	PASS
175	174	5	9	7.6	1807	PASS
176	174	95	101	98.4	23370	PASS
177	176	5	9	6.7	1571	PASS