

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091620\
 Data File : VX018438.D
 Acq On : 16 Sep 2020 14:34
 Operator : JC/SP
 Sample : L4008-02DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 13703DL

Quant Time: Sep 17 07:56:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	471951	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	728426	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	690855	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	354229	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	323193	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
35) Dibromofluoromethane	5.46	113	252440	50.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.78%	
50) Toluene-d8	8.70	98	900852	48.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.50%	
62) 4-Bromofluorobenzene	11.12	95	338510	46.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.12%	
Target Compounds						
					Qvalue	
16) Acetone	2.42	43	307004	120.252	ug/l	96
25) 2-Butanone	4.64	43	60860	14.505	ug/l	95
51) 4-Methyl-2-Pentanone	8.62	43	91117	11.704	ug/l	100
52) Toluene	8.76	92	170490	13.619	ug/l	97
67) Ethyl Benzene	10.24	91	33982	1.448	ug/l	96
68) m/p-Xylenes	10.34	106	66412	7.583	ug/l	93
69) o-Xylene	10.68	106	26067	3.070	ug/l	93
70) Styrene	10.70	104	10119	0.700	ug/l #	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091620\
Data File : VX018438.D
Acq On : 16 Sep 2020 14:34
Operator : JC/SP
Sample : L4008-02DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
13703DL

Quant Time: Sep 17 07:56:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

