

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091620\
 Data File : VX018440.D
 Acq On : 16 Sep 2020 15:19
 Operator : JC/SP
 Sample : L4009-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 2020-09-11-TB01

Quant Time: Sep 17 04:55:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	440527	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	703784	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	678615	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	336603	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	319186	52.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.38%	
35) Dibromofluoromethane	5.47	113	242919	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
50) Toluene-d8	8.70	98	842990	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	
62) 4-Bromofluorobenzene	11.12	95	328383	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
Target Compounds						
16) Acetone	2.42	43	66957	28.097	ug/l	95
25) 2-Butanone	4.64	43	18953	4.839	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091620\
Data File : VX018440.D
Acq On : 16 Sep 2020 15:19
Operator : JC/SP
Sample : L4009-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
2020-09-11-TB01

Quant Time: Sep 17 04:55:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

