

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091620\
 Data File : VX018444.D
 Acq On : 16 Sep 2020 16:50
 Operator : JC/SP
 Sample : L3988-27
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FC-DUP06-20200910

Quant Time: Sep 17 08:07:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	435441	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	698365	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	677249	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	355443	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	309461	51.19	ug/l	0.00
Spiked Amount			Recovery	=	102.38%	
35) Dibromofluoromethane	5.46	113	238910	50.23	ug/l	0.00
Spiked Amount			Recovery	=	100.46%	
50) Toluene-d8	8.70	98	861212	48.62	ug/l	0.00
Spiked Amount			Recovery	=	97.24%	
62) 4-Bromofluorobenzene	11.12	95	332355	47.69	ug/l	0.00
Spiked Amount			Recovery	=	95.38%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
6) Chloroethane	1.72	64	4866	1.628	ug/l	94
16) Acetone	2.42	43	5470	2.322	ug/l	94
24) 1,1-Dichloroethane	3.67	63	11433	1.331	ug/l	94
65) Chlorobenzene	10.12	112	126255	9.334	ug/l	99
67) Ethyl Benzene	10.24	91	4879	0.212	ug/l	92
73) Isopropylbenzene	11.00	105	29589	1.262	ug/l	100
78) n-propylbenzene	11.34	91	13188	0.495	ug/l	95
83) tert-Butylbenzene	11.76	119	19208	1.010	ug/l	80
85) sec-Butylbenzene	11.93	105	29328	1.356	ug/l	92
87) 1,3-Dichlorobenzene	12.01	146	222350	19.541	ug/l	99
88) 1,4-Dichlorobenzene	12.08	146	746569	65.530	ug/l	98
91) 1,2-Dichlorobenzene	12.38	146	195515	18.254	ug/l	98
93) 1,2,4-Trichlorobenzene	13.63	180	40632	5.689	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091620\
Data File : VX018444.D
Acq On : 16 Sep 2020 16:50
Operator : JC/SP
Sample : L3988-27
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FC-DUP06-20200910

Quant Time: Sep 17 08:07:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

