

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
 Data File : VX024277.D
 Acq On : 16 Sep 2021 10:31
 Operator : JC/MD
 Sample : VX0916WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0916WBL01

Quant Time: Sep 17 04:38:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	151820	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	261904	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	247058	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87028	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	119737	55.308	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 110.620%	
35) Dibromofluoromethane	5.391	113	93200	53.062	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.120%	
50) Toluene-d8	8.653	98	336289	52.556	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.120%	
62) 4-Bromofluorobenzene	11.085	95	119904	48.152	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.300%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
Data File : VX024277.D
Acq On : 16 Sep 2021 10:31
Operator : JC/MD
Sample : VX0916WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0916WBL01

Quant Time: Sep 17 04:38:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

