

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
 Data File : VX024284.D
 Acq On : 16 Sep 2021 13:37
 Operator : JC/MD
 Sample : M3733-09ME
 Misc : 7.09g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NAPL-01-(8-10)-END-POINTME

Quant Time: Sep 17 04:39:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	142217	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	238634	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	213508	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101274	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101825	50.21	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 100.42%		
35) Dibromofluoromethane	5.397	113	74028	46.26	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 92.52%		
50) Toluene-d8	8.653	98	285749	49.01	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 98.02%		
62) 4-Bromofluorobenzene	11.085	95	94785	41.78	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 83.56%		
Target Compounds						
27) cis-1,2-Dichloroethene	4.495	96	26795	16.16	ug/l	Qvalue 94
44) Trichloroethene	7.129	130	12698	7.18	ug/l	97
64) Tetrachloroethene	9.275	164	53793	32.52	ug/l	94
85) sec-Butylbenzene	11.896	105	9414	1.22	ug/l	99
89) n-Butylbenzene	12.335	91	6592	1.18	ug/l	# 70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
Data File : VX024284.D
Acq On : 16 Sep 2021 13:37
Operator : JC/MD
Sample : M3733-09ME
Misc : 7.09g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NAPL-01-(8-10)-END-POINTME

Quant Time: Sep 17 04:39:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

