

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
 Data File : VX024285.D
 Acq On : 16 Sep 2021 14:00
 Operator : JC/MD
 Sample : M3733-10ME
 Misc : 7.29g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NAPL-02-(8-10)-END-POINTME

Quant Time: Sep 17 04:39:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	131766	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	221947	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	223306	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98894	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	93669	49.85	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.70%
35) Dibromofluoromethane	5.404	113	69252	46.53	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.06%
50) Toluene-d8	8.653	98	274805	50.68	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.36%
62) 4-Bromofluorobenzene	11.086	95	112153	53.15	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.30%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.380	62	5733	4.36	ug/l	91
27) cis-1,2-Dichloroethene	4.489	96	380465	247.69	ug/l	92
44) Trichloroethene	7.129	130	61663	37.50	ug/l	98
64) Tetrachloroethene	9.275	164	120872	69.87	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
Data File : VX024285.D
Acq On : 16 Sep 2021 14:00
Operator : JC/MD
Sample : M3733-10ME
Misc : 7.29g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NAPL-02-(8-10)-END-POINTME

Quant Time: Sep 17 04:39:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

