

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
 Data File : VX024286.D
 Acq On : 16 Sep 2021 14:24
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Sep 17 04:40:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	140905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	244903	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232415	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101749	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104848	52.182	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.360%
35) Dibromofluoromethane	5.397	113	79302	48.284	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.560%
50) Toluene-d8	8.653	98	300855	50.283	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.560%
62) 4-Bromofluorobenzene	11.085	95	112003	48.101	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.200%
Target Compounds						
						Qvalue
20) Methylene Chloride	2.794	84	3683	2.239	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
Data File : VX024286.D
Acq On : 16 Sep 2021 14:24
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Sep 17 04:40:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

