

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
 Data File : VX024287.D
 Acq On : 16 Sep 2021 14:48
 Operator : JC/MD
 Sample : M3733-10MEDL 10X
 Misc : 7.29g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NAPL-02-(8-10)-END-POINTMEDL

Quant Time: Sep 17 04:40:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	143251	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	243359	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240062	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108498	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104537	51.176	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.360%	
35) Dibromofluoromethane	5.397	113	78944	48.370	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.740%	
50) Toluene-d8	8.653	98	305452	51.375	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.740%	
62) 4-Bromofluorobenzene	11.085	95	117862	50.939	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.880%	
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.501	96	39353	23.565	ug/l	95
44) Trichloroethene	7.129	130	6650	3.688	ug/l	96
64) Tetrachloroethene	9.275	164	12758	6.860	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
Data File : VX024287.D
Acq On : 16 Sep 2021 14:48
Operator : JC/MD
Sample : M3733-10MEDL 10X
Misc : 7.29g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NAPL-02-(8-10)-END-POINTMEDL

Quant Time: Sep 17 04:40:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

