

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
 Data File : VX024291.D
 Acq On : 16 Sep 2021 16:29
 Operator : JC/MD
 Sample : M3733-02ME
 Misc : 5.33g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SUP-UST-01-(10-12)ME

Quant Time: Sep 17 04:41:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	143281	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	242060	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	214999	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109369	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	98797	48.36	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.72%
35) Dibromofluoromethane	5.397	113	73689	45.39	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.78%
50) Toluene-d8	8.653	98	292980	49.54	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.08%
62) 4-Bromofluorobenzene	11.085	95	118251	51.38	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.76%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.782	84	14188	8.48	ug/l	94
39) Methylcyclohexane	7.379	83	5910	2.22	ug/l	89
73) Isopropylbenzene	10.970	105	31784	3.90	ug/l	98
78) n-propylbenzene	11.311	91	118820	12.91	ug/l	99
80) 1,3,5-Trimethylbenzene	11.457	105	195714	28.51	ug/l	100
83) tert-Butylbenzene	11.719	119	19877	3.00	ug/l	87
84) 1,2,4-Trimethylbenzene	11.756	105	506263	74.97	ug/l	100
85) sec-Butylbenzene	11.896	105	91027	10.96	ug/l	94
86) p-Isopropyltoluene	12.012	119	70178	10.07	ug/l	98
89) n-Butylbenzene	12.335	91	89667	14.87	ug/l #	46

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
Data File : VX024291.D
Acq On : 16 Sep 2021 16:29
Operator : JC/MD
Sample : M3733-02ME
Misc : 5.33g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SUP-UST-01-(10-12)ME

Quant Time: Sep 17 04:41:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

