

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
 Data File : VX024292.D
 Acq On : 16 Sep 2021 16:52
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Sep 17 04:41:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	146218	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	256195	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240973	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92849	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107555	51.585	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.160%
35) Dibromofluoromethane	5.398	113	81922	47.680	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.360%
50) Toluene-d8	8.653	98	308488	49.286	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.580%
62) 4-Bromofluorobenzene	11.086	95	99544	40.866	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	81.740%#
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.794	84	1789	1.048	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
Data File : VX024292.D
Acq On : 16 Sep 2021 16:52
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Sep 17 04:41:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

