

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
 Data File : VX024293.D
 Acq On : 16 Sep 2021 17:16
 Operator : JC/MD
 Sample : M3774-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-RR-02-(1.0)

Quant Time: Sep 17 04:41:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	142736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	247584	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	237653	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104404	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	106069	52.113	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 104.220%			
35) Dibromofluoromethane	5.397	113	80996	48.781	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.560%			
50) Toluene-d8	8.653	98	304411	50.326	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.660%			
62) 4-Bromofluorobenzene	11.085	95	114236	48.529	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.060%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	4565	5.260	ug/l	91
30) Chloroform	5.111	83	3352	1.143	ug/l	91
40) Benzene	6.050	78	5220	0.800	ug/l #	81
52) Toluene	8.726	92	1654	0.387	ug/l	97
65) Chlorobenzene	10.079	112	8264	1.690	ug/l	98
91) 1,2-Dichlorobenzene	12.341	146	2890	0.842	ug/l	92
95) Naphthalene	13.780	128	30790	5.307	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
Data File : VX024293.D
Acq On : 16 Sep 2021 17:16
Operator : JC/MD
Sample : M3774-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-RR-02-(1.0)

Quant Time: Sep 17 04:41:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

