

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX091621\  
 Data File : VX024295.D  
 Acq On : 16 Sep 2021 18:03  
 Operator : JC/MD  
 Sample : M3774-05  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 SW-CB-02-(0.2)

Quant Time: Sep 17 04:42:14 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091421W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Sep 14 12:18:04 2021  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	5.562	168	138265	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	242156	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229566	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97005	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104645	53.076	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	106.160%	
35) Dibromofluoromethane	5.397	113	79034	48.666	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.340%	
50) Toluene-d8	8.653	98	301783	51.010	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	102.020%	
62) 4-Bromofluorobenzene	11.085	95	109284	47.466	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	94.940%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	4250	5.055	ug/l	91
30) Chloroform	5.099	83	1313	0.462	ug/l #	81
65) Chlorobenzene	10.086	112	1464	0.310	ug/l #	66
95) Naphthalene	13.780	128	1307	1.471	ug/l #	91
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX091621\  
 Data File : VX024295.D  
 Acq On : 16 Sep 2021 18:03  
 Operator : JC/MD  
 Sample : M3774-05  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 SW-CB-02-(0.2)

Quant Time: Sep 17 04:42:14 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091421W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Sep 14 12:18:04 2021  
 Response via : Initial Calibration

