

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
 Data File : VX024297.D
 Acq On : 16 Sep 2021 18:50
 Operator : JC/MD
 Sample : M3774-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-CB-03-(1.0)

Manual Integrations
 APPROVED

MMDadoda
 9/17/2021 2:32:18 PM

Quant Time: Sep 17 04:42:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	136419	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	235443	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	209344	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84919	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	101052	51.947	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 103.900%			
35) Dibromofluoromethane	5.397	113	78064	49.439	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.880%			
50) Toluene-d8	8.653	98	295581	51.386	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 102.780%			
62) 4-Bromofluorobenzene	11.085	95	97505	43.557	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 87.120%			
Target Compounds						
8) Diethyl Ether	2.142	74	1249	1.579	ug/l	79
16) Acetone	2.386	43	4224	5.092	ug/l #	87
30) Chloroform	5.111	83	998	0.356	ug/l #	75
40) Benzene	6.050	78	125943	20.286	ug/l	98
65) Chlorobenzene	10.086	112	80592	18.712	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	1384	0.486	ug/l	96
88) 1,4-Dichlorobenzene	12.043	146	4033m	1.388	ug/l	
91) 1,2-Dichlorobenzene	12.335	146	4141	1.483	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
Data File : VX024297.D
Acq On : 16 Sep 2021 18:50
Operator : JC/MD
Sample : M3774-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
SW-CB-03-(1.0)

Manual Integrations
APPROVED

MMDadoda
9/17/2021 2:32:18 PM

Quant Time: Sep 17 04:42:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

