

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
 Data File : VX024298.D
 Acq On : 16 Sep 2021 19:14
 Operator : JC/MD
 Sample : M3774-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Sep 17 04:42:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	124672	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	217927	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209939	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102846	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	90966	51.168	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.340%
35) Dibromofluoromethane	5.397	113	72938	49.906	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.820%
50) Toluene-d8	8.653	98	265655	49.895	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.800%
62) 4-Bromofluorobenzene	11.085	95	113288	54.675	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091621\
 Data File : VX024298.D
 Acq On : 16 Sep 2021 19:14
 Operator : JC/MD
 Sample : M3774-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampled :
 TRIP-BLANK

Quant Time: Sep 17 04:42:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

