

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
 Data File : VX031408.D
 Acq On : 16 Sep 2022 14:23
 Operator : JC/MD
 Sample : VX0916WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0916WBL02

Quant Time: Sep 17 02:04:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	86965	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	145324	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	131024	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58242	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	93642	43.869	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.740%
35) Dibromofluoromethane	5.385	113	78929	50.527	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.060%
50) Toluene-d8	8.647	98	298011	50.857	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.720%
62) 4-Bromofluorobenzene	11.079	95	98187	45.758	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
Data File : VX031408.D
Acq On : 16 Sep 2022 14:23
Operator : JC/MD
Sample : VX0916WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0916WBL02

Quant Time: Sep 17 02:04:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

