

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
 Data File : VX031419.D
 Acq On : 16 Sep 2022 18:46
 Operator : JC/MD
 Sample : N4645-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOLL-MW-02

Quant Time: Sep 17 02:07:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	83048	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	134729	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125308	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58025	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82147	40.299	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	80.600%
35) Dibromofluoromethane	5.391	113	68677	47.421	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.840%
50) Toluene-d8	8.653	98	263763	48.552	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.100%
62) 4-Bromofluorobenzene	11.085	95	86882	43.673	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.340%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1325	1.523	ug/l #	83
31) Cyclohexane	5.477	56	12904	5.892	ug/l	96
39) Methylcyclohexane	7.379	83	3055	1.399	ug/l	99
40) Benzene	6.044	78	2854	0.501	ug/l	97
73) Isopropylbenzene	10.964	105	1183	0.276	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
Data File : VX031419.D
Acq On : 16 Sep 2022 18:46
Operator : JC/MD
Sample : N4645-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOLL-MW-02

Quant Time: Sep 17 02:07:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

