

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
 Data File : VX031422.D
 Acq On : 16 Sep 2022 19:56
 Operator : JC/MD
 Sample : N4645-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOLL-MW-05

Quant Time: Sep 17 02:08:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	80278	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	134874	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57009	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83649	42.452	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.900%
35) Dibromofluoromethane	5.385	113	70241	48.449	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.900%
50) Toluene-d8	8.653	98	268433	49.359	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.720%
62) 4-Bromofluorobenzene	11.079	95	87754	44.064	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.120%
Target Compounds						
19) Methyl tert-butyl Ether	3.117	73	5060	0.975	ug/l	Qvalue # 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
Data File : VX031422.D
Acq On : 16 Sep 2022 19:56
Operator : JC/MD
Sample : N4645-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOLL-MW-05

Quant Time: Sep 17 02:08:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

