

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
 Data File : VX031423.D
 Acq On : 16 Sep 2022 20:19
 Operator : JC/MD
 Sample : N4645-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOLL-MW-06

Quant Time: Sep 17 02:08:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	80515	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	132918	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118379	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53822	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80183	40.573	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	81.140%
35) Dibromofluoromethane	5.391	113	68407	47.878	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.760%
50) Toluene-d8	8.653	98	257796	48.101	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.200%
62) 4-Bromofluorobenzene	11.085	95	84057	42.829	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.660%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1686	1.998	ug/l #	85
19) Methyl tert-butyl Ether	3.117	73	14955	2.873	ug/l	97
22) Diisopropyl ether	3.763	45	1710	0.329	ug/l #	81
27) cis-1,2-Dichloroethene	4.495	96	1397	0.844	ug/l	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
Data File : VX031423.D
Acq On : 16 Sep 2022 20:19
Operator : JC/MD
Sample : N4645-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOLL-MW-06

Quant Time: Sep 17 02:08:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

