

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
 Data File : VX031424.D
 Acq On : 16 Sep 2022 20:42
 Operator : JC/MD
 Sample : N4645-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOLL-MW-07

Quant Time: Sep 17 02:09:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	81358	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	131685	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119075	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54175	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81566	40.845	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	81.700%		
35) Dibromofluoromethane	5.391	113	68720	48.548	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.100%		
50) Toluene-d8	8.653	98	260040	48.973	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.940%		
62) 4-Bromofluorobenzene	11.079	95	84060	43.232	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	86.460%		
Target Compounds						
19) Methyl tert-butyl Ether	3.117	73	12900	2.452	ug/l	94
22) Diisopropyl ether	3.757	45	2369	0.451	ug/l #	82
27) cis-1,2-Dichloroethene	4.495	96	1636	0.978	ug/l	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
Data File : VX031424.D
Acq On : 16 Sep 2022 20:42
Operator : JC/MD
Sample : N4645-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOLL-MW-07

Quant Time: Sep 17 02:09:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

