

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
 Data File : VX031431.D
 Acq On : 16 Sep 2022 23:49
 Operator : JC/MD
 Sample : PB147603TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB147603TB

Quant Time: Sep 17 04:27:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	82212	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	137232	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127430	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53231	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	49745	24.652	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	49.300%#		
35) Dibromofluoromethane	5.385	113	41887	28.395	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	56.800%#		
50) Toluene-d8	8.653	98	166142	30.025	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	60.040%#		
62) 4-Bromofluorobenzene	11.079	95	51215	25.275	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	50.540%#		
Target Compounds					Qvalue	
16) Acetone	2.386	43	6191	7.187	ug/l	96
37) Ethyl Acetate	4.721	43	13105	4.939	ug/l	98
43) Isopropyl Acetate	6.342	43	57929	12.790	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
Data File : VX031431.D
Acq On : 16 Sep 2022 23:49
Operator : JC/MD
Sample : PB147603TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB147603TB

Quant Time: Sep 17 04:27:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

