

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
 Data File : VX031436.D
 Acq On : 17 Sep 2022 01:46
 Operator : JC/MD
 Sample : PB147603ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB147603ZHE#05

Quant Time: Sep 17 04:29:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	82155	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	137250	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	133305	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57543	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	50314	24.951	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	49.900%#
35) Dibromofluoromethane	5.385	113	42579	28.860	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	57.720%#
50) Toluene-d8	8.653	98	167801	30.321	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	60.640%#
62) 4-Bromofluorobenzene	11.079	95	53876	26.585	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	53.160%#
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	6066	7.046	ug/l	100
37) Ethyl Acetate	4.715	43	12883	4.855	ug/l	96
43) Isopropyl Acetate	6.342	43	56664	12.509	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
Data File : VX031436.D
Acq On : 17 Sep 2022 01:46
Operator : JC/MD
Sample : PB147603ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB147603ZHE#05

Quant Time: Sep 17 04:29:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

