

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
 Data File : VX031438.D
 Acq On : 17 Sep 2022 02:32
 Operator : JC/MD
 Sample : PB147603ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB147603ZHE#07

Quant Time: Sep 17 04:30:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	82622	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	137727	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	136212	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59287	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	50934	25.116	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	50.240%#
35) Dibromofluoromethane	5.385	113	43261	29.221	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	58.440%#
50) Toluene-d8	8.653	98	167346	30.134	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	60.260%#
62) 4-Bromofluorobenzene	11.079	95	54239	26.671	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	53.340%#
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	8550	9.876	ug/l	96
37) Ethyl Acetate	4.721	43	13050	4.901	ug/l	96
43) Isopropyl Acetate	6.342	43	59051	12.991	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
Data File : VX031438.D
Acq On : 17 Sep 2022 02:32
Operator : JC/MD
Sample : PB147603ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB147603ZHE#07

Quant Time: Sep 17 04:30:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

