

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
 Data File : VX031441.D
 Acq On : 17 Sep 2022 03:42
 Operator : JC/MD
 Sample : PB147603ZHE#10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB147603ZHE#10

Quant Time: Sep 17 04:30:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	81394	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	135279	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	133104	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58125	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	49737	24.895	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	49.800%#
35) Dibromofluoromethane	5.391	113	42221	29.035	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	58.060%#
50) Toluene-d8	8.653	98	165818	30.399	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	60.800%#
62) 4-Bromofluorobenzene	11.079	95	52532	26.299	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	52.600%#
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	8310	9.743	ug/l #	86
37) Ethyl Acetate	4.721	43	11935	4.563	ug/l	98
43) Isopropyl Acetate	6.342	43	58079	13.009	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
Data File : VX031441.D
Acq On : 17 Sep 2022 03:42
Operator : JC/MD
Sample : PB147603ZHE#10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB147603ZHE#10

Quant Time: Sep 17 04:30:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

