

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
 Data File : VX031442.D
 Acq On : 17 Sep 2022 04:05
 Operator : JC/MD
 Sample : PB147603ZHE#11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB147603ZHE#11

Quant Time: Sep 17 04:57:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	82928	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	137630	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	135050	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59096	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51075	25.092	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	50.180%#
35) Dibromofluoromethane	5.391	113	43112	29.141	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	58.280%#
50) Toluene-d8	8.653	98	169861	30.608	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	61.220%#
62) 4-Bromofluorobenzene	11.085	95	53196	26.177	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	52.360%#
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	8435	9.707	ug/l	97
37) Ethyl Acetate	4.720	43	12393	4.657	ug/l	99
43) Isopropyl Acetate	6.342	43	57705	12.704	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
Data File : VX031442.D
Acq On : 17 Sep 2022 04:05
Operator : JC/MD
Sample : PB147603ZHE#11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB147603ZHE#11

Quant Time: Sep 17 04:57:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

