

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
 Data File : VX031443.D
 Acq On : 17 Sep 2022 04:29
 Operator : JC/MD
 Sample : PB147603ZHE#12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB147603ZHE#12

Quant Time: Sep 17 04:58:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	82159	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	139757	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	134162	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59631	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51279	25.428	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	50.860%#
35) Dibromofluoromethane	5.391	113	42761	28.464	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	56.920%#
50) Toluene-d8	8.653	98	169914	30.152	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	60.300%#
62) 4-Bromofluorobenzene	11.079	95	54526	26.423	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	52.840%#
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	8541	9.921	ug/l	99
37) Ethyl Acetate	4.721	43	12130	4.489	ug/l	99
43) Isopropyl Acetate	6.342	43	57056	12.370	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
Data File : VX031443.D
Acq On : 17 Sep 2022 04:29
Operator : JC/MD
Sample : PB147603ZHE#12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB147603ZHE#12

Quant Time: Sep 17 04:58:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

