

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
 Data File : VX031444.D
 Acq On : 17 Sep 2022 04:52
 Operator : JC/MD
 Sample : PB147603ZHE#13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB147603ZHE#13

Quant Time: Sep 17 05:35:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	80737	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	136143	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	130506	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56354	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	49837	25.148	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	50.300%#		
35) Dibromofluoromethane	5.391	113	42073	28.749	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	57.500%#		
50) Toluene-d8	8.652	98	165550	30.157	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	60.320%#		
62) 4-Bromofluorobenzene	11.079	95	51870	25.803	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	51.600%#		
Target Compounds					Qvalue	
16) Acetone	2.385	43	8341	9.859	ug/l	99
37) Ethyl Acetate	4.726	43	12061	4.582	ug/l	97
43) Isopropyl Acetate	6.342	43	58690	13.062	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
Data File : VX031444.D
Acq On : 17 Sep 2022 04:52
Operator : JC/MD
Sample : PB147603ZHE#13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB147603ZHE#13

Quant Time: Sep 17 05:35:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

