

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
 Data File : VX031445.D
 Acq On : 17 Sep 2022 05:15
 Operator : JC/MD
 Sample : PB147603ZHE#14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB147603ZHE#14

Quant Time: Sep 17 05:38:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	79013	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	133507	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	132391	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57547	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	49423	25.484	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	50.960%#
35) Dibromofluoromethane	5.391	113	41162	28.682	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	57.360%#
50) Toluene-d8	8.653	98	163208	30.318	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	60.640%#
62) 4-Bromofluorobenzene	11.079	95	53034	26.903	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	53.800%#
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	8135	9.826	ug/l #	85
37) Ethyl Acetate	4.727	43	12124	4.697	ug/l	99
43) Isopropyl Acetate	6.342	43	57526	13.056	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
Data File : VX031445.D
Acq On : 17 Sep 2022 05:15
Operator : JC/MD
Sample : PB147603ZHE#14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB147603ZHE#14

Quant Time: Sep 17 05:38:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration

