


#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 4.904 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:128 Resp: 14810
Ion Ratio Lower Upper
128 100
49 164.5 0.0 358.3
130 125.3 0.0 319.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#2
Dichlorodifluoromethane
Concen: 74.799 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26


Tgt Ion: 85 Resp: 75781
Ion Ratio Lower Upper
85 100
87 33.5 15.6 46.8





#3
Chloromethane
Concen: 98.360 ug/l
RT: 1.288 min Scan# 33
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 50 Resp: 103917
Ion Ratio Lower Upper
50 100
52 32.4 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#4
Vinyl Chloride
Concen: 102.645 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26


Tgt Ion: 62 Resp: 122882
Ion Ratio Lower Upper
62 100
64 33.2 26.2 39.2





#5
Bromomethane
Concen: 339.169 ug/l
RT: 1.612 min Scan# 86
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
Client Sample Id :
VSTDICC100




Tgt Ion: 94 Resp: 167289
Ion Ratio Lower Upper
94 100
96 94.3 77.0 115.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022




#6
Chloroethane
Concen: 208.310 ug/l
RT: 1.685 min Scan# 98
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Tgt Ion: 64 Resp: 133550
Ion Ratio Lower Upper
64 100
66 32.6 26.4 39.6





#7
 Trichlorofluoromethane
 Concen: 206.207 ug/l
 RT: 1.886 min Scan# 131
 Delta R.T. 0.000 min
 Lab File: VX031402.D
 Acq: 16 Sep 2022 11:26




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100




Tgt Ion:101 Resp: 312925
 Ion Ratio Lower Upper
 101 100
 103 64.5 49.6 74.4

Manual Integrations
 APPROVED



Reviewed By :John Carlone 09/17/2022
 Supervised By :Mahesh Dadoda 09/19/2022




#8
 Diethyl Ether
 Concen: 161.833 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. 0.000 min
 Lab File: VX031402.D
 Acq: 16 Sep 2022 11:26



Tgt Ion: 74 Resp: 104456
 Ion Ratio Lower Upper
 74 100
 45 81.4 18.7 168.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 98.557 ug/l
RT: 2.325 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100

Tgt Ion:101 Resp: 100700
Ion Ratio Lower Upper
101 100
85 48.0 0.0 84.6
151 75.6 0.0 165.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#10
1,1-Dichloroethene
Concen: 92.324 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26


Tgt Ion: 96 Resp: 96327
Ion Ratio Lower Upper
96 100
61 160.0 122.6 183.8
98 67.3 50.1 75.1





#11
Methyl Iodide
Concen: 108.227 ug/l
RT: 2.453 min Scan# 211
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:142 Resp: 126948
Ion Ratio Lower Upper
142 100
127 45.4 22.3 66.9
141 13.8 7.0 20.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#12
Methyl Acetate
Concen: 90.552 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26


Tgt Ion: 43 Resp: 153868
Ion Ratio Lower Upper
43 100
74 25.2 22.1 33.1





#13
Acrolein
Concen: 506.453 ug/l
RT: 2.240 min Scan# 189
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion: 56 Resp: 101098
Ion Ratio Lower Upper
56 100
55 70.5 56.4 84.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022




#14
Acrylonitrile
Concen: 435.587 ug/l
RT: 3.069 min Scan# 325
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Tgt Ion: 53 Resp: 318555
Ion Ratio Lower Upper
53 100
52 82.4 41.5 124.5
51 36.8 18.1 54.4





#15
Acetone
Concen: 413.752 ug/l
RT: 2.392 min Scan# 213
Delta R.T. 0.006 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 58 Resp: 82193
Ion Ratio Lower Upper
58 100
43 325.5 250.2 375.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#16
Carbon Disulfide
Concen: 93.954 ug/l
RT: 2.508 min Scan# 233
Delta R.T. -0.006 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Tgt Ion: 76 Resp: 255899
Ion Ratio Lower Upper
76 100
78 9.2 7.1 10.7






#17
 Allyl chloride
 Concen: 88.395 ug/l
 RT: 2.666 min Scan# 259
 Delta R.T. 0.000 min
 Lab File: VX031402.D
 Acq: 16 Sep 2022 11:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100






Tgt Ion: 41 Resp: 162530
 Ion Ratio Lower Upper
 41 100
 39 83.6 35.4 106.1
 76 39.7 18.4 55.4


Manual Integrations

APPROVED



Reviewed By :John Carlone 09/17/2022
 Supervised By :Mahesh Dadoda 09/19/2022




#18
 Methylene Chloride
 Concen: 90.472 ug/l
 RT: 2.788 min Scan# 279
 Delta R.T. 0.000 min
 Lab File: VX031402.D
 Acq: 16 Sep 2022 11:26




Tgt Ion: 84 Resp: 109974
 Ion Ratio Lower Upper
 84 100
 49 124.9 94.2 141.2
 51 37.5 29.2 43.8
 86 64.5 51.8 77.6





#19
trans-1,2-Dichloroethene
Concen: 103.297 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion: 96 Resp: 11695
Ion Ratio Lower Upper
96 100
61 125.8 108.7 163.1
98 62.7 49.2 73.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022




#20
Diisopropyl ether
Concen: 111.114 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26



Tgt Ion: 45 Resp: 409276
Ion Ratio Lower Upper
45 100
43 76.3 70.3 105.5
87 31.7 24.5 36.7
59 12.0 9.4 14.2









#21
1,1-Dichloroethane
Concen: 101.229 ug/l
RT: 3.611 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100

Tgt Ion: 63 Resp: 20073
Ion Ratio Lower Upper
63 100
98 7.6 3.8 11.2
100 5.7 2.5 7.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#22
cis-1,2-Dichloroethene
Concen: 98.266 ug/l
RT: 4.489 min Scan# 558
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Tgt Ion: 96 Resp: 132114
Ion Ratio Lower Upper
96 100
61 141.4 107.4 161.2
98 64.0 51.1 76.7









#23
tert-Butyl Alcohol
Concen: 438.292 ug/l m
RT: 3.008 min Scan# 312
Delta R.T. 0.031 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 59 Resp: 146840
Ion Ratio Lower Upper
59 100
52 178.7 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#24
Methyl tert-Butyl Ether
Concen: 113.495 ug/l
RT: 3.117 min Scan# 333
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Tgt Ion: 73 Resp: 382017
Ion Ratio Lower Upper
73 100
43 22.1 16.0 24.0









#25
Chloroform
Concen: 117.821 ug/l
RT: 5.099 min Scan# 657
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100

Tgt Ion: 83 Resp: 232845
Ion Ratio Lower Upper
83 100
85 64.5 49.9 74.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#26
Cyclohexane
Concen: 98.978 ug/l
RT: 5.471 min Scan# 719
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26


Tgt Ion: 56 Resp: 181847
Ion Ratio Lower Upper
56 100
89 0.0 0.0 0.0
84 96.0 76.2 114.2





#27
1,2-Dichloroethane-d4
Concen: 33.167 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion: 65 Resp: 40070
Ion Ratio Lower Upper
65 100
67 50.4 42.3 63.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022




#28
1,4-Difluorobenzene
Concen: 30.000 ug/l
RT: 6.763 min Scan# 931
Delta R.T. -0.006 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26



Tgt Ion: 114 Resp: 79331
Ion Ratio Lower Upper
114 100
63 19.3 15.3 22.9
88 15.3 12.2 18.4





#29

1,1-Dichloropropene

Concen: 122.022 ug/l

RT: 5.690 min Scan# 756

Delta R.T. 0.000 min

Lab File: VX031402.D

Acq: 16 Sep 2022 11:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100





Tgt Ion: 75	Resp: 16559
Ion Ratio	Lower Upper
75 100	
110 36.4	0.0 73.6
77 30.7	0.0 62.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/17/2022

Supervised By :Mahesh Dadoda 09/19/2022



#30

2-Butanone

Concen: 499.489 ug/l




RT: 4.568 min Scan# 571


Delta R.T. 0.006 min

Lab File: VX031402.D

Acq: 16 Sep 2022 11:26


Tgt Ion: 43	Resp: 457532
Ion Ratio	Lower Upper
43 100	
57 8.2	6.6 10.0
72 27.8	22.1 33.1





#31
2,2-Dichloropropane
Concen: 143.969 ug/l
RT: 4.477 min Scan# 557
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion: 77 Resp: 18673
Ion Ratio Lower Upper
77 100
97 22.4 18.4 27.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022




#32
1,1,1-Trichloroethane
Concen: 141.861 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Tgt Ion: 97 Resp: 207008
Ion Ratio Lower Upper
97 100
99 64.0 51.8 77.8
61 45.2 33.4 50.0





#33
Carbon Tetrachloride
Concen: 150.989 ug/l
RT: 5.678 min Scan# 753
Delta R.T. -0.006 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100



Tgt Ion: 117 Resp: 18505
Ion Ratio Lower Upper
117 100
119 96.0 75.3 112.9
121 28.7 24.1 36.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#34
Benzene
Concen: 110.167 ug/l
RT: 6.044 min Scan# 813
Delta R.T. 0.006 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26


Tgt Ion: 78 Resp: 480449
Ion Ratio Lower Upper
78 100
77 24.0 18.5 27.7
51 17.0 12.9 19.3





#35
Methacrylonitrile
Concen: 104.301 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:	41	Resp:	9670
Ion Ratio	100	Lower	Upper
41	100		
39	57.4	27.0	81.0
67	73.7	33.3	99.8
52	29.7	12.9	38.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#36
1,2-Dichloroethane
Concen: 138.885 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Tgt Ion:	62	Resp:	184380
Ion Ratio	100	Lower	Upper
62	100		
98	8.9	7.4	11.0









#37
Trichloroethene
Concen: 111.212 ug/l
RT: 7.129 min Scan# 991
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion:130 Resp: 129712
Ion Ratio Lower Upper
130 100
95 96.9 76.5 114.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#38
Methylcyclohexane
Concen: 109.246 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Tgt Ion: 83 Resp: 192374
Ion Ratio Lower Upper
83 100
55 71.4 37.0 110.9
98 50.2 25.9 77.8





#39

1,2-Dichloropropane

Concen: 105.586 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX031402.D

Acq: 16 Sep 2022 11:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 63 Resp: 11721

Ion Ratio Lower Upper

63 100





65 30.6 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/17/2022

Supervised By :Mahesh Dadoda 09/19/2022



#40

Dibromomethane

Concen: 118.646 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX031402.D

Acq: 16 Sep 2022 11:26




Tgt Ion: 93 Resp: 89282


Ion Ratio Lower Upper

93 100

95 83.2 0.0 168.8

174 95.6 0.0 220.8









#41
Bromodichloromethane
Concen: 133.343 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 83 Resp: 17924
Ion Ratio Lower Upper
83 100
85 64.1 50.6 75.8
127 8.7 7.0 10.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#42
Vinyl Acetate
Concen: 606.447 ug/l
RT: 3.727 min Scan# 433
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26


Tgt Ion: 43 Resp: 1776475
Ion Ratio Lower Upper
43 100
86 11.0 8.5 12.7





#43
Ethyl Acetate
Concen: 109.632 ug/l
RT: 4.715 min Scan# 595
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion: 43 Resp: 189164
Ion Ratio Lower Upper
43 100
45 14.7 11.8 17.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022




#44
Isopropyl Acetate
Concen: 117.692 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Tgt Ion: 43 Resp: 308497
Ion Ratio Lower Upper
43 100
61 18.8 17.0 25.4





#45
1,4-Dioxane
Concen: 2050.255 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.018 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 88 Resp: 63088
Ion Ratio Lower Upper
88 100
43 31.9 25.8 38.6
58 61.7 54.8 82.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#46
Methyl methacrylate
Concen: 111.357 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26


Tgt Ion: 41 Resp: 138325
Ion Ratio Lower Upper
41 100
69 86.2 41.9 125.8
39 71.3 27.5 82.5





#47
n-amy1 Acetate
Concen: 132.948 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. 0.006 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100



Tgt Ion: 43 Resp: 29937
Ion Ratio Lower Upper
43 100
55 24.7 20.4 30.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#48
t-1,3-Dichloropropene
Concen: 149.360 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Tgt Ion: 75 Resp: 221876
Ion Ratio Lower Upper
75 100
77 32.0 26.6 40.0





#49

cis-1,3-Dichloropropene

Concen: 126.656 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX031402.D

Acq: 16 Sep 2022 11:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100





Tgt Ion:	75	Resp:	211270
Ion	Ratio	Lower	Upper
75	100		
77	31.6	25.2	37.8
39	55.5	37.4	56.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/17/2022

Supervised By :Mahesh Dadoda 09/19/2022



#50

1,1,2-Trichloroethane

Concen: 129.086 ug/l




RT: 9.153 min Scan# 1323


Delta R.T. 0.000 min

Lab File: VX031402.D

Acq: 16 Sep 2022 11:26


Tgt Ion:	97	Resp:	143910
Ion	Ratio	Lower	Upper
97	100		
83	86.5	66.2	99.4
85	53.8	40.4	60.6
99	62.2	51.1	76.7





#51
Ethyl methacrylate
Concen: 127.884 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100




Tgt Ion: 69 Resp: 22185
Ion Ratio Lower Upper
69 100
41 64.6 32.1 96.5
39 44.1 18.3 54.9

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022




#52
1,3-Dichloropropane
Concen: 124.632 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26



Tgt Ion: 76 Resp: 223502
Ion Ratio Lower Upper
76 100
78 32.7 0.0 64.4









#53
Dibromochloromethane
Concen: 168.335 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100

Tgt Ion:129 Resp: 173700
Ion Ratio Lower Upper
129 100
127 78.1 38.4 115.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#54
1,2-Dibromoethane
Concen: 146.707 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Tgt Ion:107 Resp: 169842
Ion Ratio Lower Upper
107 100
109 94.8 76.2 114.4









#55
2-Chloroethyl vinyl ether
Concen: 637.280 ug/l
RT: 8.245 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 63 Resp: 56990
Ion Ratio Lower Upper
63 100
65 32.4 25.9 38.9
106 31.1 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#56
Bromoform
Concen: 191.933 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26


Tgt Ion:173 Resp: 138641
Ion Ratio Lower Upper
173 100
175 48.2 40.3 60.5
254 8.5 7.4 11.2





#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 117 Resp: 85148
Ion Ratio Lower Upper
117 100
82 53.8 42.3 63.5
119 32.3 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#58
4-Methyl-2-Pentanone
Concen: 493.897 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. 0.006 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Tgt Ion: 43 Resp: 1038780
Ion Ratio Lower Upper
43 100
58 40.8 31.8 47.6
85 20.7 15.8 23.6









#59
2-Hexanone
Concen: 488.921 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 79535
Ion Ratio Lower Upper
43 100
58 56.8 44.1 66.1
57 20.1 16.2 24.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#60
4-Bromofluorobenzene
Concen: 36.118 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26


Tgt Ion: 95 Resp: 57957
Ion Ratio Lower Upper
95 100
174 80.7 65.4 98.2
176 78.8 63.4 95.0





#61
Tetrachloroethene
Concen: 105.053 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion:164 Resp: 134308
Ion Ratio Lower Upper
164 100
166 125.6 105.0 157.6
129 93.2 78.5 117.7
131 88.3 71.4 107.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022




#62
Toluene
Concen: 115.439 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26



Tgt Ion: 91 Resp: 653858
Ion Ratio Lower Upper
91 100
92 58.2 48.1 72.1









#63
Toluene-d8
Concen: 30.330 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 98 Resp: 14003
Ion Ratio Lower Upper
98 100
100 65.5 51.4 77.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#64
Chlorobenzene
Concen: 117.421 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26


Tgt Ion:112 Resp: 397195
Ion Ratio Lower Upper
112 100
114 32.8 26.0 39.0





#65
1,1,1,2-Tetrachloroethane
Concen: 126.964 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:131 Resp: 14976
Ion Ratio Lower Upper
131 100
133 94.3 0.0 197.2
119 63.5 0.0 127.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#66
Ethyl Benzene
Concen: 122.963 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26


Tgt Ion: 91 Resp: 740137
Ion Ratio Lower Upper
91 100
106 32.3 25.8 38.8





#67
m/p-Xylenes
Concen: 288.366 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion:106 Resp: 673544
Ion Ratio Lower Upper
106 100
91 196.5 159.5 239.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022




#68
o-Xylene
Concen: 135.350 ug/l
RT: 10.647 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26



Tgt Ion:106 Resp: 314815
Ion Ratio Lower Upper
106 100
91 207.6 105.4 316.2





#69

Styrene

Concen: 148.049 ug/l

RT: 10.659 min Scan# 1570

Delta R.T. 0.000 min

Lab File: VX031402.D

Acq: 16 Sep 2022 11:26

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion:104 Resp: 569850
 Ion Ratio Lower Upper
 104 100
 78 47.9 39.7 59.5
 103 53.4 42.7 64.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/17/2022
 Supervised By :Mahesh Dadoda 09/19/2022



#70

Isopropylbenzene

Concen: 146.620 ug/l




RT: 10.964 min Scan# 1620


Delta R.T. 0.000 min

Lab File: VX031402.D

Acq: 16 Sep 2022 11:26


Tgt Ion:105 Resp: 858635
 Ion Ratio Lower Upper
 105 100
 120 27.1 18.8 34.8





#71
1,1,2,2-Tetrachloroethane
Concen: 117.125 ug/l
RT: 11.214 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion: 83 Resp: 249430
Ion Ratio Lower Upper
83 100
131 10.9 5.8 17.3
85 64.3 32.0 96.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022




#72
1,2,3-Trichloropropane
Concen: 133.445 ug/l m
RT: 11.244 min Scan# 1666
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26



Tgt Ion: 75 Resp: 255450
Ion Ratio Lower Upper
75 100
77 34.6 0.0 78.6









#73
Bromobenzene
Concen: 143.164 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100

Tgt Ion:156 Resp: 20104
Ion Ratio Lower Upper
156 100
77 148.5 0.0 297.6
158 97.2 0.0 188.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#74
n-propylbenzene
Concen: 141.778 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26


Tgt Ion: 91 Resp: 995998
Ion Ratio Lower Upper
91 100
120 24.1 0.0 48.0





#75
2-Chlorotoluene
Concen: 141.917 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100




Tgt Ion: 91 Resp: 587090
Ion Ratio Lower Upper
91 100
126 35.6 0.0 68.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022




#76
1,3,5-Trimethylbenzene
Concen: 158.721 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26



Tgt Ion:105 Resp: 780404
Ion Ratio Lower Upper
105 100
120 50.4 0.0 98.0









#77
t-1,4-Dichloro-2-butene
Concen: 132.526 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100

Tgt Ion: 75 Resp: 87258
Ion Ratio Lower Upper
75 100
53 92.3 59.7 179.1
89 47.7 23.1 69.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#78
4-Chlorotoluene
Concen: 154.021 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26


Tgt Ion: 91 Resp: 715117
Ion Ratio Lower Upper
91 100
126 32.4 0.0 59.8





#79
tert-butylbenzene
Concen: 149.256 ug/l
RT: 11.720 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion:119 Resp: 713928
Ion Ratio Lower Upper
119 100
91 58.8 0.0 119.4
134 24.8 0.0 48.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022




#80
1,2,4-Trimethylbenzene
Concen: 147.914 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.006 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26



Tgt Ion:105 Resp: 747678
Ion Ratio Lower Upper
105 100
120 47.2 0.0 89.6









#81
 sec-Butylbenzene
 Concen: 142.592 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX031402.D
 Acq: 16 Sep 2022 11:26

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100

Tgt Ion:105 Resp: 923214
 Ion Ratio Lower Upper
 105 100
 134 21.1 0.0 41.2




Manual Integrations
 APPROVED


Reviewed By :John Carlone 09/17/2022
 Supervised By :Mahesh Dadoda 09/19/2022



#82
 p-Isopropyltoluene
 Concen: 154.936 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX031402.D
 Acq: 16 Sep 2022 11:26


Tgt Ion:119 Resp: 807571
 Ion Ratio Lower Upper
 119 100
 134 27.2 0.0 52.4
 91 24.3 0.0 47.2





#83
1,3-Dichlorobenzene
Concen: 151.154 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100




Tgt Ion:146 Resp: 410674
Ion Ratio Lower Upper
146 100
111 39.9 19.8 59.4
148 63.2 31.6 94.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022




#84
1,4-Dichlorobenzene
Concen: 156.587 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26



Tgt Ion:146 Resp: 431468
Ion Ratio Lower Upper
146 100
111 38.2 19.2 57.6
148 64.4 32.3 96.8









#85
n-Butylbenzene
Concen: 135.124 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 91 Resp: 660920
Ion Ratio Lower Upper
91 100
92 54.4 0.0 109.4
134 28.5 0.0 56.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#86
Hexachloroethane
Concen: 153.364 ug/l
RT: 12.543 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Tgt Ion:117 Resp: 132811
Ion Ratio Lower Upper
117 100
201 68.8 39.5 118.5









#87
1,2-Dichlorobenzene
Concen: 151.139 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion:146 Resp: 40548
Ion Ratio Lower Upper
146 100
111 41.6 21.1 63.3
148 64.2 31.4 94.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#88
1,2-Dibromo-3-Chloropropane
Concen: 114.506 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Tgt Ion: 75 Resp: 55611
Ion Ratio Lower Upper
75 100
155 84.7 74.3 111.5
157 106.9 97.3 145.9









#89
1,2,4-Trichlorobenzene
Concen: 123.964 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion:180 Resp: 20591
Ion Ratio Lower Upper
180 100
182 92.2 77.4 116.2
145 31.5 24.2 36.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#90
Hexachlorobutadiene
Concen: 112.278 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26


Tgt Ion:225 Resp: 70731
Ion Ratio Lower Upper
225 100
223 64.7 0.0 126.0
227 64.7 0.0 125.0





#91
Naphthalene
Concen: 114.862 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.006 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:128 Resp: 760499
Ion Ratio Lower Upper
128 100
127 13.2 10.0 15.0
129 11.1 8.8 13.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/17/2022
Supervised By :Mahesh Dadoda 09/19/2022



#92
1,2,3-Trichlorobenzene
Concen: 123.260 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX031402.D
Acq: 16 Sep 2022 11:26

Tgt Ion:180 Resp: 208553
Ion Ratio Lower Upper
180 100
182 94.7 0.0 189.4
145 33.0 0.0 67.2

