

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
 Data File : VX031437.D
 Acq On : 17 Sep 2022 02:09
 Operator : JC/MD
 Sample : PB147603ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB147603ZHE#06

Quant Time: Sep 17 04:29:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 01:50:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.556	168	81532	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	138922	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	136423	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60187	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	50323	25.146	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	50.300%#
35) Dibromofluoromethane	5.391	113	42646	28.558	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	57.120%#
50) Toluene-d8	8.653	98	170139	30.373	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	60.740%#
62) 4-Bromofluorobenzene	11.079	95	54821	26.725	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	53.460%#
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	8511	9.962	ug/l	100
37) Ethyl Acetate	4.721	43	12489	4.650	ug/l	98
43) Isopropyl Acetate	6.342	43	58075	12.667	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091622\
Data File : VX031437.D
Acq On : 17 Sep 2022 02:09
Operator : JC/MD
Sample : PB147603ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB147603ZHE#06

Quant Time: Sep 17 04:29:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091422W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 01:50:07 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 713
Delta R.T. -0.006 min
Lab File: VX031437.D
Acq: 17 Sep 2022 02:09




Instrument :
MSVOA_X
ClientSampleId :
PB147603ZHE#06


Tgt Ion:168 Resp: 81532
Ion Ratio Lower Upper
168 100
99 54.4 48.8 73.2



#16
Acetone
Concen: 9.962 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX031437.D
Acq: 17 Sep 2022 02:09

Tgt Ion: 43 Resp: 8511
Ion Ratio Lower Upper
43 100
58 29.7 23.8 35.8









#33
 1,2-Dichloroethane-d4
 Concen: 25.146 ug/l
 RT: 5.958 min Scan# 799
 Delta R.T. 0.000 min
 Lab File: VX031437.D
 Acq: 17 Sep 2022 02:09




Instrument :
 MSVOA_X
 ClientSampleId :
 PB147603ZHE#06


Tgt Ion: 65 Resp: 50323
 Ion Ratio Lower Upper
 65 100
 67 51.8 0.0 102.4



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.763 min Scan# 931
 Delta R.T. 0.000 min
 Lab File: VX031437.D
 Acq: 17 Sep 2022 02:09

Tgt Ion: 114 Resp: 138922
 Ion Ratio Lower Upper
 114 100
 63 18.8 0.0 42.6
 88 14.0 0.0 33.0





#35

Dibromofluoromethane

Concen: 28.558 ug/l

RT: 5.391 min Scan# 705

Delta R.T. 0.006 min

Lab File: VX031437.D


Acq: 17 Sep 2022 02:09

Instrument :

MSVOA_X

ClientSampleId :

PB147603ZHE#06






Tgt Ion:113 Resp: 42646

Ion Ratio Lower Upper

113 100

111 101.4 82.4 123.6

192 18.1 14.1 21.1



#37

Ethyl Acetate

Concen: 4.650 ug/l

RT: 4.721 min Scan# 596

Delta R.T. 0.000 min

Lab File: VX031437.D

Acq: 17 Sep 2022 02:09




Tgt Ion: 43 Resp: 12489


Ion Ratio Lower Upper

43 100

61 13.5 11.1 16.7

70 9.3 8.6 13.0









#43
Isopropyl Acetate
Concen: 12.667 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.000 min
Lab File: VX031437.D
Acq: 17 Sep 2022 02:09




Instrument :
MSVOA_X
ClientSampleId :
PB147603ZHE#06


Tgt Ion: 43 Resp: 58075
Ion Ratio Lower Upper
43 100
61 18.3 15.3 22.9
87 13.4 10.6 15.8



#50
Toluene-d8
Concen: 30.373 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. 0.000 min
Lab File: VX031437.D
Acq: 17 Sep 2022 02:09

Tgt Ion: 98 Resp: 170139
Ion Ratio Lower Upper
98 100
100 64.4 52.5 78.7









#62
4-Bromofluorobenzene
Concen: 26.725 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX031437.D
Acq: 17 Sep 2022 02:09




Instrument :
MSVOA_X
ClientSampleId :
PB147603ZHE#06


Tgt Ion: 95 Resp: 54821
Ion Ratio Lower Upper
95 100
174 88.2 0.0 153.2
176 85.3 0.0 144.0



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX031437.D
Acq: 17 Sep 2022 02:09

Tgt Ion: 117 Resp: 136423
Ion Ratio Lower Upper
117 100
82 52.8 46.8 70.2
119 31.9 26.4 39.6





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX031437.D


Acq: 17 Sep 2022 02:09

Instrument :

MSVOA_X

ClientSampleId :

PB147603ZHE#06



Tgt Ion:152 Resp: 60187

Ion Ratio Lower Upper

152 100

115 56.1 43.4 130.1

150 156.2 0.0 346.8

