

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091625\
 Data File : VX047593.D
 Acq On : 16 Sep 2025 14:03
 Operator : JC/MD
 Sample : VX0916MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0916MBL01

Quant Time: Sep 17 06:48:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	133117	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	268042	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	260832	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	128834	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	113017	53.338	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.680%
35) Dibromofluoromethane	5.367	113	91020	50.633	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.260%
50) Toluene-d8	8.629	98	326402	52.475	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.940%
62) 4-Bromofluorobenzene	11.061	95	139939	59.279	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	118.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091625\
Data File : VX047593.D
Acq On : 16 Sep 2025 14:03
Operator : JC/MD
Sample : VX0916MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0916MBL01

Quant Time: Sep 17 06:48:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

