

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091625\
 Data File : VX047598.D
 Acq On : 16 Sep 2025 16:30
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Sep 17 06:53:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	192687	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	334451	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	297936	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	148925	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	150627	49.111	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.220%
35) Dibromofluoromethane	5.373	113	112834	50.305	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.600%
50) Toluene-d8	8.635	98	409153	52.717	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.440%
62) 4-Bromofluorobenzene	11.061	95	156102	52.996	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091625\
Data File : VX047598.D
Acq On : 16 Sep 2025 16:30
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Sep 17 06:53:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

