

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091718\
 Data File : VX004629.D
 Acq On : 17 Sep 2018 19:58
 Operator : JC/MD
 Sample : PB112894ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB112894ZHE#04

Quant Time: Sep 18 01:30:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	220336	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	324329	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	312414	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	178726	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	146426	51.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.08%	
35) Dibromofluoromethane	5.51	113	140502	47.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.68%	
50) Toluene-d8	8.71	98	457686	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
62) 4-Bromofluorobenzene	11.14	95	156911	44.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.34%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	18042	13.01	ug/l	94
18) Methyl Acetate	2.78	43	7904	2.03	ug/l	99
20) Methylene Chloride	2.85	84	10941	3.85	ug/l	89
25) 2-Butanone	4.70	43	4911	2.23	ug/l #	89
43) Isopropyl Acetate	6.46	43	147554	24.81	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091718\
Data File : VX004629.D
Acq On : 17 Sep 2018 19:58
Operator : JC/MD
Sample : PB112894ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB112894ZHE#04

Quant Time: Sep 18 01:30:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

