

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091718\
 Data File : VX004633.D
 Acq On : 17 Sep 2018 21:45
 Operator : JC/MD
 Sample : PB112894ZHE#08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB112894ZHE#08

Quant Time: Sep 18 01:46:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	227984	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	314291	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	302167	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	178241	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	141442	48.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.24%	
35) Dibromofluoromethane	5.49	113	133275	46.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.68%	
50) Toluene-d8	8.71	98	441342	45.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.52%	
62) 4-Bromofluorobenzene	11.14	95	151066	43.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.76%	
Target Compounds						
16) Acetone	2.44	43	18785	13.09	ug/l	96
18) Methyl Acetate	2.78	43	9869	2.45	ug/l	99
20) Methylene Chloride	2.86	84	12527	4.26	ug/l	92
43) Isopropyl Acetate	6.46	43	147247	25.55	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091718\
Data File : VX004633.D
Acq On : 17 Sep 2018 21:45
Operator : JC/MD
Sample : PB112894ZHE#08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB112894ZHE#08

Quant Time: Sep 18 01:46:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

