

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091718\
 Data File : VX004638.D
 Acq On : 17 Sep 2018 23:59
 Operator : JC/MD
 Sample : PB112894ZHE#13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB112894ZHE#13

Quant Time: Sep 18 02:11:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	223362	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	318376	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	310523	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	191917	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	151475	52.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.18%	
35) Dibromofluoromethane	5.51	113	142080	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
50) Toluene-d8	8.71	98	447252	45.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.56%	
62) 4-Bromofluorobenzene	11.14	95	156423	44.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.70%	
Target Compounds						
16) Acetone	2.44	43	23508	16.72	ug/l	95
18) Methyl Acetate	2.78	43	10295	2.61	ug/l	97
20) Methylene Chloride	2.85	84	9469	3.28	ug/l #	86
25) 2-Butanone	4.70	43	6262	2.80	ug/l #	89
43) Isopropyl Acetate	6.46	43	160393	27.47	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091718\
Data File : VX004638.D
Acq On : 17 Sep 2018 23:59
Operator : JC/MD
Sample : PB112894ZHE#13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB112894ZHE#13

Quant Time: Sep 18 02:11:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration

