


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX091718\
Data File : VX004620.D
Acq On : 17 Sep 2018 14:23
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Title : SW846 8260
Last Update : Wed Sep 12 02:17:23 2018



AutoFind: Scans 1637, 1638, 1639; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	1530	PASS
75	95	30	60	48.3	4388	PASS
95	95	100	100	100.0	9092	PASS
96	95	5	9	6.3	576	PASS
173	174	0.00	2	1.3	110	PASS
174	95	50	100	95.8	8708	PASS
175	174	5	9	7.1	621	PASS
176	174	95	101	98.5	8576	PASS
177	176	5	9	6.7	576	PASS