

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091719\
 Data File : VX012437.D
 Acq On : 17 Sep 2019 17:21
 Operator : JC/SP
 Sample : VX0917WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0917WBL01

Quant Time: Sep 18 03:42:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	144938	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	232146	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	202692	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	93612	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	100621	48.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.88%	
35) Dibromofluoromethane	5.48	113	70333	50.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.12%	
50) Toluene-d8	8.71	98	271723	52.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.50%	
62) 4-Bromofluorobenzene	11.13	95	108544	50.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.54%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091719\
Data File : VX012437.D
Acq On : 17 Sep 2019 17:21
Operator : JC/SP
Sample : VX0917WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0917WBL01

Quant Time: Sep 18 03:42:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

