

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091719\
 Data File : VX012440.D
 Acq On : 17 Sep 2019 18:42
 Operator : JC/SP
 Sample : K4858-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20190909

Quant Time: Sep 18 03:51:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	131113	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	207369	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	180856	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	73734	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	89108	47.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.84%	
35) Dibromofluoromethane	5.49	113	62865	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
50) Toluene-d8	8.71	98	242437	52.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.36%	
62) 4-Bromofluorobenzene	11.14	95	89945	46.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.26%	
Target Compounds						
16) Acetone	2.43	43	15301	10.184	ug/l	98
20) Methylene Chloride	2.86	84	2598	2.284	ug/l #	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091719\
Data File : VX012440.D
Acq On : 17 Sep 2019 18:42
Operator : JC/SP
Sample : K4858-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-20190909

Quant Time: Sep 18 03:51:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

