

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091719\
 Data File : VX012441.D
 Acq On : 17 Sep 2019 19:05
 Operator : JC/SP
 Sample : K4858-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D3-20190909

Quant Time: Sep 18 04:02:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	125863	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	202567	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	166076	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	60077	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	87932	48.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.50%	
35) Dibromofluoromethane	5.49	113	62348	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
50) Toluene-d8	8.71	98	231380	50.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.98%	
62) 4-Bromofluorobenzene	11.14	95	77070	40.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.80%	
Target Compounds						
44) Trichloroethene	7.21	130	16092	15.626	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091719\
Data File : VX012441.D
Acq On : 17 Sep 2019 19:05
Operator : JC/SP
Sample : K4858-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D3-20190909

Quant Time: Sep 18 04:02:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

