

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091719\
 Data File : VX012442.D
 Acq On : 17 Sep 2019 19:29
 Operator : JC/SP
 Sample : K4858-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW202D-20190909

Quant Time: Sep 18 04:05:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	134709	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	214853	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	177287	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	67778	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	92747	48.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.08%	
35) Dibromofluoromethane	5.49	113	66051	50.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.58%	
50) Toluene-d8	8.71	98	247081	51.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.66%	
62) 4-Bromofluorobenzene	11.14	95	87146	43.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.20%	
Target Compounds						
					Qvalue	
7) Trichlorofluoromethane	1.92	101	5571	3.797	ug/l	95
12) 1,1-Dichloroethene	2.36	96	2368	2.796	ug/l	82
24) 1,1-Dichloroethane	3.69	63	15519	7.026	ug/l	98
27) cis-1,2-Dichloroethene	4.58	96	19965	15.196	ug/l	88
30) Chloroform	5.21	83	2430	0.958	ug/l	98
32) 1,1,1-Trichloroethane	5.49	97	2291	1.081	ug/l #	44
42) 1,2-Dichloroethane	6.18	62	4129	2.144	ug/l	82
44) Trichloroethene	7.20	130	114528	104.853	ug/l	93
64) Tetrachloroethene	9.33	164	28062	34.542	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091719\
Data File : VX012442.D
Acq On : 17 Sep 2019 19:29
Operator : JC/SP
Sample : K4858-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW202D-20190909

Quant Time: Sep 18 04:05:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

