

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091719\
 Data File : VX012444.D
 Acq On : 17 Sep 2019 20:15
 Operator : JC/SP
 Sample : K4858-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE135D1-20190910

Quant Time: Sep 18 04:09:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	129782	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	206112	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	170309	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	62360	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	89659	48.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.40%	
35) Dibromofluoromethane	5.49	113	63403	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
50) Toluene-d8	8.71	98	237392	51.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.82%	
62) 4-Bromofluorobenzene	11.14	95	81551	42.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.08%	
Target Compounds						
30) Chloroform	5.21	83	4113	1.683	ug/l	Qvalue 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091719\
Data File : VX012444.D
Acq On : 17 Sep 2019 20:15
Operator : JC/SP
Sample : K4858-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE135D1-20190910

Quant Time: Sep 18 04:09:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

