

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091719\
 Data File : VX012452.D
 Acq On : 17 Sep 2019 23:24
 Operator : JC/SP
 Sample : K4858-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D1-20190909

Quant Time: Sep 18 04:51:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	162948	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	261697	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	220631	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	92236	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	113696	48.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.36%	
35) Dibromofluoromethane	5.49	113	79929	50.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.92%	
50) Toluene-d8	8.71	98	303958	51.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.68%	
62) 4-Bromofluorobenzene	11.14	95	111241	45.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.40%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.38	101	6138	4.780	ug/l	98
12) 1,1-Dichloroethene	2.37	96	1543	1.506	ug/l	96
27) cis-1,2-Dichloroethene	4.59	96	4619	2.906	ug/l	88
38) Carbon Tetrachloride	5.77	117	1868	0.944	ug/l #	72
44) Trichloroethene	7.21	130	1087423	817.351	ug/l	96
49) 1,4-Dioxane	7.75	88	1021	20.310	ug/l #	85
64) Tetrachloroethene	9.34	164	4424	4.376	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091719\
Data File : VX012452.D
Acq On : 17 Sep 2019 23:24
Operator : JC/SP
Sample : K4858-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D1-20190909

Quant Time: Sep 18 04:51:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

