

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091719\
 Data File : VX012456.D
 Acq On : 18 Sep 2019 00:57
 Operator : JC/SP
 Sample : K4858-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D1-20190910

Quant Time: Sep 18 05:09:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	152331	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	243440	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	212663	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	86649	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	105043	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery =	96.22%		
35) Dibromofluoromethane	5.49	113	74763	50.74	ug/l	0.00
Spiked Amount 50.000			Recovery =	101.48%		
50) Toluene-d8	8.71	98	286848	52.60	ug/l	0.00
Spiked Amount 50.000			Recovery =	105.20%		
62) 4-Bromofluorobenzene	11.14	95	107329	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery =	94.80%		

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.37	101	8015	6.676	ug/l	96
12) 1,1-Dichloroethene	2.37	96	1702	1.777	ug/l #	76
19) Methyl tert-butyl Ether	3.19	73	4003	0.767	ug/l #	82
27) cis-1,2-Dichloroethene	4.59	96	11267	7.584	ug/l	84
30) Chloroform	5.20	83	7339	2.558	ug/l	97
44) Trichloroethene	7.21	130	444105	358.842	ug/l	94
49) 1,4-Dioxane	7.74	88	1828	39.089	ug/l #	79
64) Tetrachloroethene	9.33	164	25512	26.180	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091719\
Data File : VX012456.D
Acq On : 18 Sep 2019 00:57
Operator : JC/SP
Sample : K4858-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D1-20190910

Quant Time: Sep 18 05:09:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

