

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX091719\
 Data File : VX012457.D
 Acq On : 18 Sep 2019 01:21
 Operator : JC/SP
 Sample : K4858-17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D2-20190910

Quant Time: Sep 18 05:13:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	153201	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	250755	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	218870	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	97711	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	107577	48.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.98%	
35) Dibromofluoromethane	5.49	113	77011	50.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.48%	
50) Toluene-d8	8.71	98	294944	52.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.00%	
62) 4-Bromofluorobenzene	11.14	95	113855	48.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.62%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	359725	297.934	ug/l	99
12) 1,1-Dichloroethene	2.37	96	4278	4.442	ug/l	78
27) cis-1,2-Dichloroethene	4.59	96	12823	8.582	ug/l	86
32) 1,1,1-Trichloroethane	5.48	97	1086	0.450	ug/l #	49
44) Trichloroethene	7.21	130	223748	175.517	ug/l	98
49) 1,4-Dioxane	7.73	88	1989	41.292	ug/l	97
64) Tetrachloroethene	9.33	164	22670	22.603	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091719\
Data File : VX012457.D
Acq On : 18 Sep 2019 01:21
Operator : JC/SP
Sample : K4858-17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D2-20190910

Quant Time: Sep 18 05:13:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

