


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX091719\  
Data File : VX012427.D  
Acq On : 17 Sep 2019 11:30  
Operator : JC/SP  
Sample : BFB  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X091719W.M  
Title : SW846 8260  
Last Update : Tue Sep 17 15:27:10 2019



AutoFind: Scans 1680, 1681, 1682; Background Corrected with Scan 1673

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 50          | 95           | 15           | 40           | 19.4      | 5727    | PASS             |
| 75          | 95           | 30           | 60           | 52.6      | 15546   | PASS             |
| 95          | 95           | 100          | 100          | 100.0     | 29552   | PASS             |
| 96          | 95           | 5            | 9            | 6.5       | 1917    | PASS             |
| 173         | 174          | 0.00         | 2            | 0.3       | 72      | PASS             |
| 174         | 95           | 50           | 100          | 70.8      | 20920   | PASS             |
| 175         | 174          | 5            | 9            | 7.6       | 1595    | PASS             |
| 176         | 174          | 95           | 101          | 95.6      | 20009   | PASS             |
| 177         | 176          | 5            | 9            | 7.4       | 1479    | PASS             |