

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091720\
 Data File : VX018474.D
 Acq On : 17 Sep 2020 18:17
 Operator : JC/SP
 Sample : L4058-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Sep 18 07:13:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	519572	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	822343	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	775120	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	385511	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	348710	47.44	ug/l	0.00
Spiked Amount			Recovery	=	94.88%	
35) Dibromofluoromethane	5.46	113	278591	49.00	ug/l	0.00
Spiked Amount			Recovery	=	98.00%	
50) Toluene-d8	8.70	98	1008302	48.82	ug/l	0.00
Spiked Amount			Recovery	=	97.64%	
62) 4-Bromofluorobenzene	11.12	95	380193	46.27	ug/l	0.00
Spiked Amount			Recovery	=	92.54%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091720\
Data File : VX018474.D
Acq On : 17 Sep 2020 18:17
Operator : JC/SP
Sample : L4058-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Sep 18 07:13:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

